BIOSPECTRIX VACCUUM TISSUE PROCESSOR: MODEL NO.: BSRD-800

ABOUT THE PRODUCT



DETAIL PRESENTATION

1. While staying in working situation, it is keeping in vacuum condition. It can avoid the reagent volatilization and waste, it decrease the cost of operation, and keep the work environment.
2. While processing the tissue, it stays at vacuum condition. While change the reagent, but the bottle needn't change. So to avoid the tissue be destroyed. No any risk while the main power cut off.
3. The crust material uses the stainless steel. There model (3.8L/5.3L/6.8L) can choose.
4. Adopt large LCD intelligent display, friendly interface, can know the working situation clearly.
5. It can set the start time automatically within one month.
6. While the tissue processing finished, it has alarm to remind customer.
7. It have 8 sets program, can offer customer choosing.
8. it have overheating protection.
9. It can remember the steps while the main power cut off. After the power comes back, it will go on running at the failure step.
10. Three models can choose, can processing 100, 200, or 300 tissues at same time separately.









BIOSPECTRIX PVT LTD

info@biospectrix.co

INNOVATION • QUALITY • RELIABLE

★★★★★

BioSpectrix

Model No. BSRD-800 Technical Specifications





Reagent Bottle capacity:	2.3L/3.2L/4.8L
Reagent bottle:	9pcs
Cleaning bottle:	3pcs
Paraffin tank:	3pcs
Processing cylinder capacity:	3.8L/5.3L/6.8L
Reagent temp.:	30 to 70°C
Paraffin tank Temp:	From ambient to 80°C
Dipping time:	0-99H59M
Mixing time:	1 minute
Mixing interval time:	10 minutes
Run mode:	manual or auto
Nominal voltage:	AC220V
Nominal frequency:	50HZ